

Certificate



Certificate no.

CU 72180546 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Singapore Pte. Ltd.
North Tech Lobby 3 #02-13/18
29 Woodlands Industrial Park E1
Singapore 757716
Singapore

Test report no.: USA-AS 31593844 010

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012
UL 61010-2-081:2015
CAN/CSA-C22.2 NO. 61010-1-12
CAN/CSA-C22.2 NO. 61010-2-081:15

Certified Product: Genetic Sequencer

License Fee - Units

Model Designation: MiniSeq

7

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 240W


Protection Class: I

Special Remarks: Replaces Certificate CU72153075.

7

Appendix: 1, 1-4

Licensed Test mark:



Date of Issue
(day/mo/yr)

21/03/2018